



| | | |
|---|---|----------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQ8F00AZ |
|---|---|----------|

TECHNOTHROMBIN® TGA SUB 5x1.5mL

REF 5006235

LOT 8F71B00.01

 2019-04-30

| Inhalt Content | Reagenz Reagent | LOT |  |
|-------------------|--------------------|---------|---|
| 5x1.5mL | TGA Substrat | 8F71B00 | 2019-04-30 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|---|---|----------------------------|--------------------|--------------|
| Kontrollwiederfindung Recovery of controls | TGA Control High (nM Peak Thrombin) LOT 1Z71B00 | TGA RC High LOT 7F72B00 | 336 - 560 | 470.6 OK |
| | | TGA RC Low LOT 7D62B00 | 207 - 345 | 273.6 OK |
| | TGA Control Low (nM Peak Thrombin) LOT 2863B01 | TGA RC High LOT 7F72B00 | 63 - 147 | 117.8 OK |
| | | TGA RC Low LOT 7D62B00 | 57 - 143 | 93.3 OK |

Durchführung von
Prepared by

Fikret NASUFI 

06. NOV. 2017

Datum / Date:

Freigabe von
Approved by

Michael KAFKA 

06. NOV. 2017

Datum / Date: