

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ1Z00AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA Control High 5x1mL

REF      5006320  
LOT      1Z71B00.01  
      30-Jun-2018

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x1mL	TGA Control High	1Z71B00	30-Jun-2018

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F72B00	336 - 560	455.9    OK
	TGA Reagenz RC Low (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 7D62B00	207 - 345	289.3    OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI *Fikret Nasufi*

20. SEP. 2017  
Datum / Date:

**Freigabe von**  
**Approved by**

Suzana GRESAK PODAKAR  
*Suzana Gresak Podakar*

20. SEP. 2017  
Datum / Date: