
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ3U00AZ.01
---	---	-------------

**TECHNOVIEW Arixtra Control L 6x1mL**

REF      5090012  
LOT      3U71B00.01  
      2019-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Richtwert Expected value Ceveron® alpha	Anti Xa	0.32 – 0.59 µg/mL	0.45	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa	0.32 – 0.59 µg/mL	0.37	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

*Lucas ROSENMAYR*  
\_\_\_\_\_  
Lucas ROSENMAYR

*27. OKT. 2017*  
\_\_\_\_\_  
Datum / Date:

**Freigabe von**  
**Approved by**

*Eva Maria HAGER*  
\_\_\_\_\_  
Eva Maria HAGER

*27. OKT. 2017*  
\_\_\_\_\_  
Datum / Date: