

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1000.01
---	---	-----------

TECHNOVIEW Apixaban CAL 5x1mL

REF 5090269

LOT 1071B00.01

 2018-09-30

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW Apixaban Cal 1	1162B00	2018-09-30
1x1mL	TECHNOVIEW Apixaban Cal 2	1262B00	2018-09-30
1x1mL	TECHNOVIEW Apixaban Cal 3	1362B00	2018-09-30
1x1mL	TECHNOVIEW Apixaban Cal 4	1462B00	2018-10-31
1x1mL	TECHNOVIEW Apixaban Cal 5	1562B00	2018-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Apixaban Cont low LOT 3971B00	74.9 – 139.1 ng/mL	91.2	OK
	TECHNOVIEW Apixaban Cont high LOT 3D71B00	259.5 – 432.5 ng/mL	335.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas ROSENMAYR

02. NOV. 2017

Datum / Date:

Freigabe von
Approved by


Stefan STEINER

02. NOV. 2017

Datum / Date: