
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3V00AZ.01
-----------------------------------------------------------------------------------	-------------------------------------------------------------	-------------

TECHNOVIEW Arixtra CONT H 6x1mL

REF 5090014
LOT 3V71B00.01
 2019-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Richtwert Expected value Ceveron® alpha	Anti Xa	0.59 – 0.98 µg/mL	0.78	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa	0.59 – 0.98 µg/mL	0.76	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas ROSENMAYR

27. OKT. 2017

Datum / Date:

Freigabe von
Approved by

Eva-Maria HAGER

27. OKT. 2017

Datum / Date: