
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3D00.01
---	---	-----------

TECHNOVIEW® Apixaban CONT High 5x1mL

REF 5090270
LOT 3D71B00.01
 2019-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		259.5 – 432.5 ng/mL	346.0	OK
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa Apixaban	259.5 – 432.5 ng/mL	335.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Lucas
Lucas ROSENMAYR

02. NOV. 2017

Datum / Date:

**Freigabe von
Approved by**

Stefan
Stefan STEINER

02. NOV. 2017

Datum / Date: