

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Z00AZ.02
---	---	-------------

TECHNOCHROM® FXIII 3 x 3 mL

REF	5360010
LOT	0Z74B00.02
	2018-09-30

Inhalt	Reagenz	LOT	
3 x 3 mL	FXIII Activator Reagent	2134100/1	2018-09-30
3 x 3 mL	FXIII Detection Reagent	2134100/2	2018-09-30
3 x 3 mL	FXIII NADPH	2134100/3	2018-09-30
3 x 1 mL	FXIII Inhibitor Reagent	2134100/4	2018-09-30
1 x 6 mL	Stabilizer solution	2134100/5	2018-09-30

Test	Spezifikation	Ergebnis	
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P71B00	58.2 – 97.0 %	92.5
	Coagulation Control A Lot: 3P71B00	28.8 – 48.0 %	37.7

Durchführung von
Prepared by


Alessandro BENIGNI

31. OKT. 2017

Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI

31. OKT. 2017

Datum / Date: