



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRB00AZ

TECHNOZYM® vWF:CBA ELISA Kit 96T

REF 5450301

LOT RB7BB00.01

⌚ 2019-04-30

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	vWF:CBA Plate Collagen Typ III	RNC75B0	2019-10-31
1 x 0.3mL	CBA Conjugate	RNE71B0	2019-04-30
1x 80mL	Wash buffer Concentrate	RNA72B0	2019-11-30
1 x 90mL	Incubation buffer	RNB73B0	2019-10-31
1 x 12mL	TMB Substrate	RKN71B0	2019-05-31
1 x 12mL	Stop solution	RKK72B0	2019-08-31
1 x 0.5mL	vWF Calibrator 1	1ND61B1	2019-07-31
1 x 0.5mL	vWF Calibrator 2	2ND61B1	2019-07-31
1 x 0.5mL	vWF Calibrator 3	3ND61B1	2019-07-31
1 x 0.5mL	vWF Calibrator 4	4ND61B1	2019-07-31
1 x 0.5mL	vWF Calibrator 5	5ND61B1	2019-07-31
1 x 0.5mL	vWF High Control	AND61B1	2019-07-31
1 x 0.5mL	vWF Low Control	BND61B1	2019-07-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
OD	Calibrator 1	> 1.0	1.3	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND61B1)	0.63 – 1.05 IU/mL	0.85	OK
	Low Control (BND61B1)	0.13 – 0.29 IU/mL	0.20	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

23. NOV. 2017
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *MK*

23. NOV. 2017
Datum / Date: