



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF 5450451

LOT RP78B0S.01

⌚ 2019-04-30

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 6	Adamts13 INH Plate	RPS74BS	2019-04-30
1 x 0.3mL	Adamts13 INH Conjugate	RPV71B0	2019-04-30
1x 80mL	Wash buffer Concentrate	RNA71B0	2019-04-30
1 x 90mL	Incubation buffer	RNB72B0	2019-04-30
1 x 12mL	TMB Substrate	RKN71B0	2019-05-31
1 x 12mL	Stop solution	RKK71B0	2019-05-31
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH High Control	RPY71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Low Control	RPZ71B0	2019-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.58	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY71B0)	60.24 – 100.40 U/mL	79.52	OK
	Low Control (RPZ71B0)	16.64 – 38.84 U/mL	27.80	OK


Durchführung von  
Prepared by

Michael KAFKA 

09. NOV. 2017

Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI 

09. NOV. 2017

Datum / Date: