

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	Antithrombin 1 x 25 IU	PRODUCT #:	81079639
EXP. DATE:	2019-09	LOT #:	N0972590
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	654171

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Antithrombin Activity	25	30	IU/vial	28	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0030	PASS
Intercept @ 37°C	0.40	0.60		0.52	PASS
r ²	0.990	--		1.000	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
Antithrombin, 25 IU	N0972590	2019-09	1 ml

PREPARED BY: *M. Oettinger*

DATE: 10-16-17

APPROVAL SIGNATURE: *P. M. Patel*

DATE: 10.16.17