



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ6C0BAZ.03

TECHNOCLOT® PT PLUS 5x4mL

REF 5003220

LOT 6C74BA0.01

 30-Apr-2019


Analytische Prozedur Analytical Procedure		Anforderung Requirement		Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN	PT Plus (%)	68.7 – 86.5	70.0	OK
	LOT 1P71B00	PT Plus (INR)	0.93 – 1.17	1.10	OK
	CCA	PT Plus (%)	10.9 – 13.7	11.0	OK
	LOT 3P71B00	PT Plus (INR)	2.59 – 3.27	3.17	OK
Etiketten und Barcode Überprüfung Label and bar code Verification				OK	

Durchführung von
Prepared by


Christoph ENNE

22. AUG. 2017
Datum / Date:

Freigabe von
Approved by

Suzana GRESAK PODAKAR


22. AUG. 2017
Datum / Date: