



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ7C00AZ

TECHNOTHROMBIN[®] TGA RB 5x0.5mL

REF 5006209

LOT 7C71B00.03

 2019-06-30

| Inhalt Content | Reagenz Reagent | LOT |  |
|-------------------|--------------------|---------|-------------------------------------------------------------------------------------|
| 5x0.5 mL | TGA RB | 7C71B00 | 2019-06-30 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|-----------------------------------------------|-------------------------------------------------------|----------------------------|--------------------|----|
| Kontrollwiederfindung Recovery of controls | TGA Control High (nM Peak Thrombin) LOT 1Z71B00 | 145 - 241 | 172.4 | OK |

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

07. DEZ. 2017

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

07. DEZ. 2017

Datum / Date: