



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3400.01

TECHNOVIEW® Rivaroxaban Control Low 5x1mL

REF 5090172

LOT 3472B00.01

⌚ 2019-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.2 – 7.9	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		40.8 – 75.8 ng/mL	58.3	OK
Verwendbarkeit Usability Ceveron Alpha®	TECHNOCHROM® anti-Xa	40.8 – 75.8 ng/mL	71.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Christoph ENNE

27. NOV. 2017
Datum / Date:

Freigabe von
Approved by


Eva Maria HAGER

27. NOV. 2017
Datum / Date: