
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3300.01
---	---	-----------

TECHNOVIEW® Rivaroxaban Control High 5x1mL

REF 5090174
LOT 3371B00.01
 2019-09-30


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		242.0 – 403.4 ng/mL	322.7	
Verwendbarkeit Usability Ceveron® alpha	TECHNOCHROM® anti-Xa	242.0 – 403.4 ng/mL	333.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Christoph ENNE

20. NOV. 2017
Datum / Date:

Freigabe von
Approved by

Suzana GRESAK PODAKAR


20. NOV. 2017
Datum / Date: