
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9J0B01
---	---	----------

Imidazol Puffer 25mL

REF 5410008
LOT 9J54BC0.02
 30-Sep-2020

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.2 – 7.6	7.5	OK
Verwendbarkeit Usability Ceveron® alpha Fibrinogen, Clauss Method	CCN LOT 1P37B01	2.52 – 3.78 g/L	3.38 OK
	CCA LOT 3P38B01	0.93 – 1.39 g/L	1.17 OK
Etiketten Überprüfung Label Verification		OK	

**Durchführung von
Prepared by**

Suzana GRESAK PODAKAR


27. MRZ. 2017

Datum / Date: _____

**Freigabe von
Approved by**


Stefan STEINER

27. MRZ. 2017

Datum / Date: _____