



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRK00AZ

TECHNOZYM® vWF:CBA ELISA Collagen Typ I Kit 96T

REF 5450311
LOT RK73B00.01
⌚ 2018-10-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	vWF:CBA Plate Collagen Typ I	RNM71B0	2019-03-31
1 x 0.3mL	vWF:Ag Conjugate	RNH61B0	2018-10-31
1x 80mL	Wash buffer Concentrate	RNA71B0	2019-04-30
1 x 90mL	Incubation buffer	RNB72B0	2019-04-30
1 x 12mL	TMB Substrate	RKN71B0	2019-05-31
1 x 12mL	Stop solution	RKK72B0	2019-08-31
1 x 0.5mL	vWF:CBA Calibrator 1	1ND61B1	2019-07-31
1 x 0.5mL	vWF:CBA Calibrator 2	2ND61B1	2019-07-31
1 x 0.5mL	vWF:CBA Calibrator 3	3ND61B1	2019-07-31
1 x 0.5mL	vWF:CBA Calibrator 4	4ND61B1	2019-07-31
1 x 0.5mL	vWF:CBA Calibrator 5	5ND61B1	2019-07-31
1 x 0.5mL	vWF:CBA High Control	AND61B1	2019-07-31
1 x 0.5mL	vWF:CBA Low Control	BND61B1	2019-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.57	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND61B1)	0.51 – 0.85 IU/mL	0.65	OK
	Low Control (BND61B1)	0.10 – 0.24 IU/mL	0.19	OK

Durchführung von
Prepared by

Michael KAFKA 

23. OKT. 2017
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

23. OKT. 2017
Datum / Date: