

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7PT00AZ
---	---	-----------

TECHNOZYM[®] ADAMTS-13 INH Control Set 2x0.5mL

REF 5450463
LOT 7PT72B0.01
 2019-08-31

Inhalt Content	Reagenz Reagent	LOT	
1 x 0.5mL	Adamts13 INH High Control	RPY71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Low Control	RPZ71B0	2019-08-31


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.37	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY71B0)	60.24 – 100.40 U/mL	81.06	OK
	Low Control (RPZ71B0)	16.64 – 38.84 U/mL	27.39	OK

**Durchführung von
Prepared by**

Michael KAFKA 

27. NOV. 2017
Datum / Date:

**Freigabe von
Approved by**

Fikret NASUFI 

27. NOV. 2017
Datum / Date: