
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6A0.01
---	---	----------

TECHNOCLOT® PT Owren manual 10x10 mL

REF 5005037
LOT 6M78BB0.01
 2019-11-30

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result										
Verwendbarkeit Usability Thrombotrack	CCN LOT 1P71B00 CCA LOT 3P71B00	<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="text-align: center;">91.4 - 115.2%</td> <td style="text-align: center;">107.8</td> <td rowspan="2" style="text-align: center; vertical-align: middle;">OK</td> </tr> <tr> <td style="text-align: center;">0.91 - 1.15 INR</td> <td style="text-align: center;">1.04</td> </tr> <tr> <td style="text-align: center;">23.9 - 30.1%</td> <td style="text-align: center;">25.3</td> <td rowspan="2" style="text-align: center; vertical-align: middle;">OK</td> </tr> <tr> <td style="text-align: center;">1.84 - 2.31 INR</td> <td style="text-align: center;">2.16</td> </tr> </table>	91.4 - 115.2%	107.8	OK	0.91 - 1.15 INR	1.04	23.9 - 30.1%	25.3	OK	1.84 - 2.31 INR	2.16
91.4 - 115.2%	107.8	OK										
0.91 - 1.15 INR	1.04											
23.9 - 30.1%	25.3	OK										
1.84 - 2.31 INR	2.16											
Etiketten und Barcode Überprüfung Label and bar code Verification		OK										

Durchführung von
Prepared by


Eva Maria HAGER

11. JAN. 2018
Datum / Date:

Freigabe von
Approved by


Stefan STEINER

11. JAN. 2018
Datum / Date: