
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0T00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA Kit 3x16 T.


REF 5006010

LOT 0T81B00.01

 2019-01-31

Inhalt Content	Reagenz Reagent	LOT	
1x3mL	TGA buffer	9P31B03	2019-11-30
3x1.5mL	TGA substrat	8F71B00	2019-04-30
1x1.5mL	TGA reagent RD	7G61B01	2019-01-31
1x0.5mL	TGA reagent RC Low	7D62B01	2019-01-31
1x0.5mL	TGA reagent RC High	7F72B00	2019-08-31
1x0.5mL	TGA Thrombin Calibrator	5T71B	2019-07-31
1x1mL	TGA Control High	1Z71B01	2019-01-31
1x1mL	TGA Control Low	2881B00	2019-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC High	TGA Control High (nM Peak Thrombin) LOT 1Z71B01	336 - 560	475.6	OK
Recovery of controls RC High	TGA Control Low (nM Peak Thrombin) LOT 2881B00	257 - 429	343.0	OK
Kontrollwiederfindung RC Low	TGA Control High (nM Peak Thrombin) LOT 1Z71B01	207 - 345	229.8	OK
Recovery of controls RC Low	TGA Control Low (nM Peak Thrombin) LOT 2881B00	132 - 220	158.5	OK
Kontrollwiederfindung RD	TGA Control High (nM Peak Thrombin) LOT 1Z71B01	369 - 615	609.0	OK
Recovery of controls RD	TGA Control Low (nM Peak Thrombin) LOT 2881B00	454 - 756	619.7	OK

	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQ0T00AZ</p>
---	---	---

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

19. JAN. 2018
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

19. JAN. 2018
Datum / Date: