
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1Z00AZ
---	---	----------

TECHNOTHROMBIN® TGA Control High 5x1mL

REF 5006320


LOT 1Z71B01.01

 2019-01-31

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1Z71B01	2019-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC High (nM Peak Thrombin) LOT 7F72B00	336 - 560	504.3 OK
	TGA Reagenz RC Low (nM Peak Thrombin) LOT 7D62B01	207 - 345	293.2 OK
	TGA Reagenz RD (nM Peak Thrombin) LOT 7G61B01	369 - 615	609.0 OK

**Durchführung von
Prepared by**

Fikret NASUFI 

19. JAN. 2019
Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA 

19. JAN 2019
Datum / Date: