

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ0U00.02
---	---	-----------

### AK-Calibrant 4x1 mL

REF	5010004
LOT	0U72B00.02
⌚	2019-09-30

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1mL	AK Calibrant A	1N74B00	2019-09-30
1x1mL	AK Calibrant B	3L74B00	2019-10-31
1x1mL	AK Calibrant C	3M74B00	2019-09-30
1x1mL	AK Calibrant D	3N78B00	2019-10-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P71B00 Technoplastin HIS	96.7 – 121.9 %	99.1	OK
	CCA LOT 3P71B00 Technoplastin HIS	24.2 – 30.6 %	27.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
**Stefan STEINER**

13. DEZ. 2017  
Datum / Date:

**Freigabe von**  
**Approved by**

  
**Eva Maria HAGER**

13. DEZ. 2017  
Datum / Date: