



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ6E0B.01

Fibrinogen Reagenz 5x5mL

REF 5138080

LOT 6E7ABB0.01

⌚ 2019-10-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.0 – 8.0	7.7	OK
Restfeuchte Residual moisture	≤3%	0.92	OK

Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P71B00	3.17 – 4.75 g/L	3.50	OK
	CCA LOT 3P71B00	1.02 – 1.54 g/L	1.20	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Levdin NASUFI

01. DEZ. 2017
Datum / Date:

Freigabe von
Approved by

Stefan STEINER

01. DEZ. 2017
Datum / Date: