

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0J00.01
---	---	-----------

TECHNOCHROM® C1-Inhibitor 30T.

REF 5345003
LOT 0J81B00.01
 2019-02-28

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Coagulation Control N for C1-INH	1P72BC1H	2019-02-28
1x1mL	Coagulation Control A for C1-INH	3P73BC1H	2019-03-31
1x0.982 mL	Coagulation Reference for C1-INH	1R72BC1H	2019-03-31
1x3 mL	C1 Esterase	7B71B00	2019-04-30
1x3 mL	Substrate f. C1-Esterase	8C72B00	2022-06-30
1x25 mL	C1-INH Probenpuffer A	9M62B00	2021-12-31
1x20 mL	C1-INH Reaktionspuffer B	9N62B00	2021-12-31


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P72BC1H	88.1 – 146.9%	118.7	OK
	CCA LOT 3P73BC1H	41.7 – 69.5%	55.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Eva Maria HAGER

24. JAN. 2018
Datum / Date:

**Freigabe von
Approved by**


Suzana GRESAK PODAKAR

25. JAN. 2018
Datum / Date: