

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ0Z00AZ.02
---	---	-------------

## TECHNOCHROM® FXIII 3 x 3 mL

**REF**                    5360010  
**LOT**                    0Z81B00.01  
                    2018-12-31

Inhalt	Reagenz	<b>LOT</b>	
3 x 3 mL	FXIII Activator Reagent	2217100/1	2018-12-31
3 x 3 mL	FXIII Detection Reagent	2217100/2	2018-12-31
3 x 3 mL	FXIII NADPH	2217100/3	2018-12-31
3 x 1 mL	FXIII Inhibitor Reagent	2217100/4	2018-12-31
1 x 6 mL	Stabilizer solution	2217100/5	2018-12-31

Test		Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P71B00	58.2 – 97.0 %	80.7
	Coagulation Control A Lot: 3P71B00	28.8 – 48.0 %	40.5

Durchführung von  
Prepared by

Michael KAFKA 

17. JAN. 2019  
Datum / Date:

Freigabe von  
Approved by

Filiret NASUFI 

17. JAN. 2019  
Datum / Date: