

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRP00AZ
---	---	----------

TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF 5450451
LOT RP81B0S.01
 2019-05-31

Inhalt Content	Reagenz Reagent	LOT	
1 x 6	Adamts13 INH Plate	RPS77BS	2019-08-31
1 x 0.3mL	Adamts13 INH Conjugate	RPV72B0	2019-21-31
1x 80mL	Wash buffer Concentrate	RNA72B0	2019-11-30
1 x 90mL	Incubation buffer	RNB73B0	2019-10-31
1 x 12mL	TMB Substrate	RKN71B0	2019-05-31
1 x 12mL	Stop solution	RKK72B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH High Control	RPY71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Low Control	RPZ71B0	2019-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.54	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (RPY71B0)	60.24 – 100.40 U/mL	80.44	OK
	Low Control (RPZ71B0)	16.64 – 38.84 U/mL	29.68	OK

Durchführung von
Prepared by

Michael KAFKA 

19. JAN. 2019
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

22. JAN. 2019
Datum / Date: