



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ6PT00AZ

TECHNOZYM[®] ADAMTS-13 INH Calibrator Set 5x0.5mL

REF 5450461

LOT 6PT72B0.01

⌚ 2019-08-31

| Inhalt Content | Reagenz Reagent | LOT | ⌚ |
|-------------------|---------------------------|---------|------------|
| 1 x 0.5mL | Adamts13 INH Calibrator 1 | 1PT71B0 | 2019-08-31 |
| 1 x 0.5mL | Adamts13 INH Calibrator 2 | 2PT71B0 | 2019-08-31 |
| 1 x 0.5mL | Adamts13 INH Calibrator 3 | 3PT71B0 | 2019-08-31 |
| 1 x 0.5mL | Adamts13 INH Calibrator 4 | 4PT71B0 | 2019-08-31 |
| 1 x 0.5mL | Adamts13 INH Calibrator 5 | 5PT71B0 | 2019-08-31 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|------------------------|----------------------------|--------------------|----|
| OD | Calibrator 1 | > 1.0 | 1.37 | OK |
| | Calibrator 5 | < 0.1 | 0.01 | OK |
| Verwendbarkeit Usability | High Control (RPY71B0) | 60.24 – 100.40 U/mL | 81.06 | OK |
| | Low Control (RPZ71B0) | 16.64 – 38.84 U/mL | 27.39 | OK |

Durchführung von
Prepared by

Michael KAFKA

27. NOV. 2017

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

27. NOV. 2017

Datum / Date: