



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ9600.01

Ceveron® Cleaning Solution 3% 25 mL

REF 9820300

LOT 9673B00.01

⌚ 2018-08-31

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	11.5 – 12.5	12.1	OK
Volumen Volume	25.0 – 26.5mL	26.5	OK
Sterilität Sterility	0 CFU/mL	0	OK

Verwendbarkeit, TECHNOVIEW Rivaroxaban Usability, TECHNOVIEW Rivaroxaban Ceveron® alpha	CONT M LOT 3571B00	86.2 – 143.6 ng/mL	110.4	OK
	CONT H LOT 3371B00	242.0 – 403.4 ng/mL	307.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Christoph ENNE

04. DEZ. 2017
Datum / Date:

Freigabe von
Approved by

Stefan STEINER

04. DEZ. 2017
Datum / Date: