

CHROMOGENIX

Certificate of Analysis

PRODUCT: CoaTest SP4 FVIII Kit	PRODUCT #: 82409463
EXP. DATE: 2019-11	LOT #: N1274202
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 824094

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0079	PASS
Intercept	--	0.250	--	0.085	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS
Low Range					
Slope	0.020	0.060	--	0.037	PASS
Intercept	--	0.406	--	0.111	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N1174040	2020-11	(1x) 6 ml
CT SP Substrate	N1174042	2019-11	(1x) 12 ml
Factor Reagent	N1274201	2020-06	(4x) 3 ml
CT SP Buffer, pH 7.3	N1173752	2019-11	(1x) 20 ml
CT SP Phospholipid	N1173753	2019-11	(1x) 2 ml

PREPARED BY: NPATEL 10-m-patel DATE: 12/20/17 ¹²⁻²⁰⁻¹⁷

APPROVAL SIGNATURE: DTENCZA P. Dencza DATE: 12/20/17 ^{12/20/17}