

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2222, 25 mg Substrate	PRODUCT #:	82031639
EXP. DATE:	2020-12	LOT #:	N0871732
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820316

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.0	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2222	N0871732	2020-12	6 ml

PREPARED BY: *MOettinger M Oettinger* DATE: 9/1/2017 9-1-17

APPROVAL SIGNATURE: *NPatel N-m-patel* DATE: 9/1/2017 9-1-17