CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2302, 25 mg Substrate	PRODUCT # :	00082034039
EXP. DATE:	2021-03	LOT # :	N1173918
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820340

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.3	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.0	PASS

DESCRIPTION	LOT NUMBER	EXP.DATE	VOLUME
S-2302	N1173918	2021-03	3 mL

PREPARED BY: Maria Oellinger DATE: 12-11-17

APPROVAL SIGNATURE: D. Amezga DATE: 12/11/17

Form QCM00820340-2 Rev. 03/09