
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7D00AZ
-----------------------------------------------------------------------------------	-------------------------------------------------------------	----------

## TECHNOTHROMBIN<sup>®</sup> TGA RC Low 5x0.5mL

REF            5006212

LOT            7D62B01.01

            2019-01-31

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RC Low	7D62B01	2019-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z72B00	251 - 418	326.3	OK
	TGA Control Low (nM Peak Thrombin) LOT 2881B00	132 - 220	197.4	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI *Fikret Nasufi*

02. FEB. 2018

Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA *[Signature]*

02. FEB. 2018

Datum / Date: