



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ1P00AZ.01

Coagulation Control N 5x1mL

REF 5020040

LOT 1P71B00.03

🕒 2019-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	F VIII chromogen	72.8 – 121.3 %	87.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Christoph ENNE

09. JAN. 2018

Datum / Date:

Freigabe von
Approved by

Eva Maria HAGER

09. JAN. 2018

Datum / Date: