
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3400.01
---	---	-----------

TECHNOVIEW® Rivaroxaban Control Low 5x1mL

REF 5090172
LOT 3481B00.01
 2020-02-29


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.2 – 7.9	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		40.7 – 75.5 ng/mL	58.1	OK
Verwendbarkeit Usability Ceveron Alpha®	TECHNOCHROM® anti-Xa	40.7 – 75.5 ng/mL	65.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Christoph ENNE

16. FEB. 2018
Datum / Date:

**Freigabe von
Approved by**

Suzana GRESAK PODAKAR


16. FEB. 2018
Datum / Date: