



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0F00AZ.03

F VIII Inhibitor Reagent Kit 2-4T.

REF 5152005

LOT 0F72B00.01

⌚ 2018-07-31

Inhalt Content	Reagenz Reagent	LOT	⌚
2x2.92 mL	FVIII normal Plasma	1G62B00	2018-11-30
1x1 mL	FVIII Inhibitor Plasma	2K71B00	2019-08-31
1x1 mL	FVIII Inhibitor Free Plasma	2M62B0D	2018-07-31
1x17 mL	Imidazol Puffer	9J74BA0	2022-07-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability FVIII clotting, Siron LS Ceveron® alpha	FVIII Inhibitor Plasma LOT 2K71B00	11.9 – 19.9 BE/mL	17.6 OK
	FVIII Inhibitor Free Plasma LOT 2M62B0D	< 0.4 BE/mL	< 0.4 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by

Christoph ENNE

30. NOV. 2017
Datum / Date:

Freigabe von
Approved by

Suzana GRESAK PODAKAR

30. NOV. 2017
Datum / Date: