



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ4Z00AZ.02

Factor XII Mangelplasma, nativ 5x1mL

REF 5194008

LOT 4Z71B00.01

⌚ 28-Feb-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Fibrinogengehalt Fibrinogen content		> 1.5g/L	2.04	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P71B00	89.3 – 148.9%	103.1	OK
	CCA LOT 3P71B00	43.1 – 71.8%	50.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Christoph ENNE

01. AUG. 2017  
Datum / Date:

Freigabe von  
Approved by

Stefan STEINER

01. AUG. 2017  
Datum / Date: