
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ6ND00AZ
---	---	-----------


**TECHNOZYM® vWF:CBA Calibrator Set 5x0.5mL**

REF      5450310  
LOT      6ND81B0.01  
      2019-12-31

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 0.5mL	vWF Calibrator 1	1ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 2	2ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 3	3ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 4	4ND71B0	2019-12-31
1 x 0.5mL	vWF Calibrator 5	5ND71B0	2019-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.66	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND71B0)	0.84 – 1.40 IU/mL	1.00	OK
	Low Control (BND71B0)	0.18 – 0.42 IU/mL	0.28	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI   
\_\_\_\_\_

31. JAN. 2019  
\_\_\_\_\_

**Freigabe von**  
**Approved by**

Michael KAFKA   
\_\_\_\_\_

31. JAN 2019  
\_\_\_\_\_