

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT: FXa, 10 x 71 nkat	PRODUCT #: 82098539
EXP. DATE: 2020-05	LOT #: N1174047
VENDOR: INSTRUMENTATION LABORATORY	ASSAYED PER QCM: 650515

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					PASS/FAIL
ACTIVITY	61	81	nkat/vial	77	PASS
STANDARD CURVE (r ²)	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	67.8	75.0	nkat/vial	69.6	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP. DATE	VOLUME
FXa, 71 nkat	N1174047	2020-05	2 ml

PREPARED BY: M. Oettinger *M. Oettinger* DATE: 12/12/2017 12-12-17

APPROVAL SIGNATURE: D. Tencza *D. Tencza* DATE: 12/12/2017 12/12/17