

# CERTIFICATE OF ANALYSIS

## CHROMOGENIX

PRODUCT: <u>COAMATIC HEPARIN</u>	PRODUCT # : <u>82339363</u>
EXP. DATE: <u>2019-09</u>	LOT # : <u>N1072744</u>
VENDOR: <u>INSTRUMENTATION LABORATORY</u>	ASSAYED PER QCM: <u>823393</u>

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
<b>A. FACTOR Xa (H01382)</b>					
pH	7.9	8.1	pH	7.9	PASS
ACTIVITY	31.5	38.5	nkat/vial	37.9	PASS
STANDARD CURVE (r <sup>2</sup> )	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	67.8	75.0	nkat/vial	69.8	PASS

<b>B. S-2732, FOR COAMATIC HEPARIN (H01391)</b>					
SUBSTRATE CONCENTRATION	14.3	15.7	mg/vial	14.7	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

<b>C. COAMATIC HEPARIN KIT (082339363) - LMW HEPARIN ON TOP</b>					
SLOPE	-0.351	-0.103	--	-0.190	PASS
r <sup>2</sup>	0.985	--	--	0.998	PASS

<b>D. COAMATIC HEPARIN KIT (082339363) - UNFRACTIONATED HEPARIN ON TOP</b>					
SLOPE	-0.336	-0.114	--	-0.185	PASS
r <sup>2</sup>	0.985	--	--	0.997	PASS

**Kit Components:**

Name	Lot Number	Exp. date	Volume
Factor Xa	N1072743	2019-10	1 ml (x 2)
S-2732	N0972595	2019-09	2 ml (x 2)

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY:

NPATEL N.M. Patel

DATE: 10/23/2017

APPROVAL SIGNATURE:

DTENCZA D. Dencza

DATE: 10/23/2017