

CERTIFICATE OF ANALYSIS



DiaPharma Group, Inc.
8948 Beckett Road, West Chester, OH 45069

Product:	DiaPharma Factor X kit
Item Number:	DPGFX
Vendors:	R ² Diagnostics, Chromogenix
Lot #:	DPGFX050
Exp. Date:	2019-07-31

Kit Components:			
Name	Lot Number	Exp. Date	Volume (1 vial each)
FXa Chromogenic Substrate	N0579844	2020-05-31	25 mg
Russell's Viper Venom	DR-1011	2019-07-31	<1 mg
CaCl ₂	DC-1011	2019-07-31	20 ml
Tris Buffer	DB-1011	2019-07-31	100 ml

Test Description	Specifications		Results		Disposition
	Min	Max	Units	Value	Pass/Fail
Slope	1.0e-3	4.0e-3	NA	3.012e-03	Pass
Intercept	-.09	.09	NA	9.573e-03	Pass
Regression Coefficient	R > 0.990		NA	0.999	Pass

Prepared By: *Wendy A. Hamill II* Date: 2018-01-25
Approved By: *[Signature]* Date: 2018-02-13