



Certificate of Analysis

Tris BSA - 100 mL

Stock Solution – 0.5 M Tris pH 7.3 (20°C), 2.0 M NaCl, 10% Bovine Serum Albumin (BSA)
Dilute 1 + 9 with water prior to use to obtain buffer working solution.

NB: The vials are slightly over dispensed. Always measure up the desired volume prior to 10-fold dilution with water.

REF	Catalogue number	TB035-100
LOT	Batch code	TTP1550-100
	Use by	2021 06
	Storage temperature	2-8°C

Analysis	Specification	Result
pH (after dilution 1 + 9 with water)	7.3 ± 0.06 at 20°C (Temp. coefficient = -0.028 per °C)	7.32 at 20°C
Conductivity (after dilution 1 + 9 with water)	20.0 ± 1 mS/cm at 20°C (Temp. coefficient = +2.1% per °C)	20.5 mS/cm at 20°C
Test of Function Chromogenic FVIII method	Approved	Approved

This Lot# of Tris BSA has been approved for release after analyses according to established protocols with results demonstrating fulfillment of product specifications.

All data concerning materials used, production data and release testing are documented and traceable under the individual Lot# which is stated on each vial.



Pia Bryngelhed
Quality Manager, Rossix AB

January 12, 2018

Date



Rossix AB
Taljegårdsgatan 3B
SE-431 53 Molndal
Sweden

Phone: +46 31 706 8965
Fax: +46 31 706 8966
info@rossix.com
www.rossix.com