

TECO medical

TECO® Human Alpha GST Serum & Urine ELISA KIT

Certificate of Analysis

Analysezertifikat

Certificat d'Analyse

REF	TE1056
LOT	051017

EXP	2018-10-31
-----	------------

Refer to Product Insert for further Information
 Weitere Informationen hierzu finden Sie in der Packungsbeilage des Produkts
 Se référer à la Notice produit pour plus d'informations

Standard Curve

Eichkurve

Courbe Standard

Standards	<u>A</u> Target Values (µg/L)	<u>X</u> Extinction 450 nm	<u>X</u> Extinction 405 nm
A	64	-	1,496
B	32	2,572	0,843
C	16	1,179	0,403
D	8	0,523	0,196
E	4	0,249	0,111
F	2	0,133	-
G	1	0,083	-
H	0	0,043	0,048
C1 - Control 1	4,4 (2,8 - 5,9)	0,273	-
C2 - Control 2	15,0 (9,8 - 20,3)	1,094	0,697

ITEM

LOT

EXP

① αGST Antibody Coated Microtiter Plate	PQ051017	2019-04-30
② Wash Buffer 50x	WA861117	2019-11-30
③ Sample Diluent	DT051017	2019-04-30
⑥ Enzyme Conjugate	EY051017	2018-10-31
⑦ TMB Substrate	TA450717	2019-08-27
⑧ Stop Solution	SA661017	2019-10-31
S Standard Stock	ZN040717	2019-04-30
C1 Control 1	ZN051017	2019-04-30
C2 Control 2	ZN051017	2019-04-30

TECOmedical

Qc release

Date

2017-11-16