

TECO medical

TECO® Human Alpha GST Serum & Urine ELISA KIT

Certificate of Analysis

Analysezertifikat

Certificat d'Analyse

REF
LOT

TE1056

061217



2018-12-31

Refer to Product Insert for further Information

Weitere Informationen hierzu finden Sie in der Packungsbeilage des Produkts

Se référer à la Notice produit pour plus d'informations

Standard Curve

Eichkurve

Courbe Standard

Standards	Δ Target Values ($\mu\text{g/L}$)	\bar{x} Extinction 450 nm	\bar{x} Extinction 405 nm
A	64	-	1,504
B	32	2,434	0,758
C	16	0,987	0,313
D	8	0,372	0,123
E	4	0,152	0,055
F	2	0,071	-
G	1	0,036	-
H	0	0,010	0,011
C1 - Control 1	5,9 (3,8 - 8,0)	0,265	-
C2 - Control 2	25,5 (16,5 - 34,4)	1,795	1,212

ITEM

LOT

EXP

① α GST Antibody Coated Microtiter Plate	PQ061217	2019-05-31
② Wash Buffer 50x	WA861117	2019-11-30
③ Sample Diluent	DT061217	2019-05-31
⑥ Enzyme Conjugate	EY061217	2018-12-31
⑦ TMB Substrate	TA450717	2019-08-27
⑧ Stop Solution	SA661017	2019-10-31
S Standard Stock	ZN051217	2019-05-31
C1 Control 1	ZN061217	2019-05-31
C2 Control 2	ZN061217	2019-05-31

TECOmedical

Qc release

Date

2018-01-08