



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRH00AZ

TECHNOZYM® D-Dimer ELISA Kit 96T

REF 2599006

LOT RH71B00.01

⌚ 2019-05-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	D-Dimer Plate	RKS71B0	2019-09-30
1x0.3mL	D-Dimer Conjugate	RKV61B1	2019-06-30
1x80mL	Wash buffer Concentrate	RNA72B0	2019-11-30
1x90mL	Incubation buffer	RNB73B0	2019-10-31
1x12mL	TMB Substrate	RKN71B0	2019-05-31
1x12mL	Stop solution	RKK72B0	2019-08-31
1x0.5mL	D-Dimer Calibrator 1	1KT31B3	2019-06-30
1x0.5mL	D-Dimer Calibrator 2	2KT31B3	2019-06-30
1x0.5mL	D-Dimer Calibrator 3	3KT31B3	2019-06-30
1x0.5mL	D-Dimer Calibrator 4	4KT31B3	2019-06-30
1x0.5mL	D-Dimer Calibrator 5	5KT42B2	2019-06-30
1x0.5mL	D-Dimer High Control	RKY31B3	2019-06-30
1x0.5mL	D-Dimer Low Control	RKZ31B3	2019-06-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	0.67	*
	Calibrator 5	< 0.1	0.04	OK
Verwendbarkeit Usability	High Control (RKY31B3)	881.4 – 1469.0 ng/mL	1078.98	OK
	Low Control (RKZ31B3)	208.1 – 346.8 ng/mL	245.29	OK

*Sonderfreigabe von RKS71B0

technoclone



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRH00AZ

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

30. NOV. 2017

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

30. NOV. 2017

Datum / Date: