



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ7G00AZ

TECHNOTHROMBIN<sup>®</sup> TGA RD Reagenz 50x1.5mL

REF 5006222

LOT 7G61B00.01

⌚ 30-Jun-2018

Inhalt Content	Reagenz Reagent	LOT	⌚
50x1.5 mL	TGA RD Reagenz	7G61B00	30-Jun-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1B52B00)	341 - 481	407.9	OK
	TGA Control Low (nM Peak Thrombin) LOT 2B61B03)	228 - 347	249.4	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

10. NOV. 2016

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

10. NOV. 2016

Datum / Date: