
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7G00AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA RD Reagenz 5x1.5mL

REF            5006220

LOT            7G71B00.01

            2020-01-31

Inhalt Content	Reagenz Reagent	LOT	
5x1.5 mL	TGA RD Reagenz	7G71B00	2020-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z72B00	432 - 720	602.0	OK
	TGA Control Low (nM Peak Thrombin) LOT 2881B00	454 - 756	572.8	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI 

26. FEB. 2018  
Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA 

27. FEB. 2018  
Datum / Date: