



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ1Z00AZ

TECHNOTHROMBIN® TGA Control High 5x1mL

REF 5006320

LOT 1Z72B00.01

⌚ 2018-12-31

Inhalt Content	Reagenz Reagent	LOT	⌚
5x1mL	TGA Control High	1Z72B00	2018-12-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC High (nM Peak Thrombin) LOT 7F72B00	375 - 625 490.7 OK
	TGA Reagenz RC Low (nM Peak Thrombin) LOT 7D81B00	251 - 418 304.4 OK
	TGA Reagenz RD (nM Peak Thrombin) LOT 7G71B00	432 - 720 514.9 OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

01. FEB. 2018

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *Michael Kafka*

01. FEB. 2018

Datum / Date: