



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ2800AZ

TECHNOTHROMBIN® TGA Control Low 5x1mL

REF 5006330

LOT 2881B00.02

⌚ 2019-01-31

Inhalt Content	Reagenz Reagent	LOT	⌚
5x1mL	TGA Control Low	2881B00	2019-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC High (nM Peak Thrombin) LOT 7F72B00	257 - 429	278.3 OK
	TGA Reagenz RC Low (nM Peak Thrombin) LOT 7D81B00	132 - 220	200.9 OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

15. MRZ. 2018

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

15. MRZ 2018

Datum / Date: