
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0D00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA Cal Set

REF 5006345
LOT 0D73B00.02
 2019-07-31

Inhalt Content	Reagenz Reagent	LOT	
1 x 3 mL	TGA buffer	9P31B03	2019-11-30
1 x 0.5mL	TGA Thrombin Calibrator	5T71B	2019-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kalibrierung / Calibration gegen / against WHO 01/580		report	722	OK
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197074		Regression der Bezugskurve / regression of the calibration curve	≥ 0.99	0.99 OK
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 2881B00	TGA RC High LOT 7F72B00	257 - 429	265.0 OK
		TGA RC Low LOT 7D81B00	132 - 220	180.0 OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

15. MRZ. 2018

Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *[Signature]*

15. MRZ. 2018

Datum / Date: