



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ1000.01

TECHNOVIEW Apixaban CAL 5x1mL

REF 5090269

LOT 1081B00.01

🕒 2020-01-31

| Inhalt<br>Content | Reagenz<br>Reagent        | LOT     | 🕒          |
|-------------------|---------------------------|---------|------------|
| 1x1mL             | TECHNOVIEW Apixaban Cal 1 | 1162B01 | 2020-01-31 |
| 1x1mL             | TECHNOVIEW Apixaban Cal 2 | 1262B01 | 2020-01-31 |
| 1x1mL             | TECHNOVIEW Apixaban Cal 3 | 1362B01 | 2020-01-31 |
| 1x1mL             | TECHNOVIEW Apixaban Cal 4 | 1462B01 | 2020-01-31 |
| 1x1mL             | TECHNOVIEW Apixaban Cal 5 | 1562B01 | 2020-01-31 |

| Analytische Prozedur<br>Analytical Procedure                         | Anforderung<br>Requirement                   | Ergebnis<br>Result  |       |    |
|--|--|---------------------|-------|----|
| Verwendbarkeit<br>Usability<br>Ceveron® alpha                        | TECHNOVIEW Apixaban Cont<br>low LOT 3961B00  | 74.1 – 137.5 ng/mL  | 95.7  | OK |
|  | TECHNOVIEW Apixaban Cont<br>high LOT 3D61B00 | 201.2 – 335.4 ng/mL | 255.4 | OK |
| Etiketten und Barcode Überprüfung<br>Label and bar code Verification |  |                     | OK    |    |

Durchführung von  
Prepared by

Suzana GRESAK PODAKAR

08. MRZ. 2018

Datum / Date:

Freigabe von  
Approved by

Christoph ENNE

08. MRZ. 2018

Datum / Date: