

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Z00AZ.02
---	---	-------------

TECHNOCHROM® FXIII 3 x 3 mL

REF 5360010
LOT 0Z82B00.01
 2019-02-28

Inhalt	Reagenz	LOT	
3 x 3 mL	FXIII Activator Reagent	2222100/1	2019-02-28
3 x 3 mL	FXIII Detection Reagent	2222100/2	2019-02-28
3 x 3 mL	FXIII NADPH	2222100/3	2019-02-28
3 x 1 mL	FXIII Inhibitor Reagent	2222100/4	2019-02-28
1 x 6 mL	Stabilizer solution	2222100/5	2019-02-28

Test		Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P71B00	58.2 – 97.0 %	83.2
	Coagulation Control A Lot: 3P71B00	28.8 – 48.0 %	38.0

Durchführung von
Prepared by

Michael KAFKA 

20. MRZ. 2018

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

20. MRZ. 2018

Datum / Date: